

Human IgM

Rabbit anti-human IgM Antibody (Polyclonal)

REFERENCES AND PRESENTATIONS¹

- **Ready-to-use (VitroStainer 42)²**
MAD-005029QDS-1
MAD-005029QDS-2
MAD-005029QDS-3
- **ready-to-use (manual or LabVision AutoStainer)**
MAD-005029QD-3
MAD-005029QD-7
MAD-005029QD-12
- **Ready-to-use (MD-Stainer)³**
MAD-005029QD-3/V
MAD-005029QD/V
- **concentrated**
MAD-005029Q - 1:50 recommended dilution

COMPOSITION

Anti-human IgM rabbit polyclonal antibody purified from serum and prepared in 10mM PBS, pH 7.4, with 0.2% BSA and 0.09% sodium azide

Intended use^{IVD}: Immunohistochemistry (IHC) on paraffin embedded tissues. Not tested on frozen tissues or Western-Blotting

Clone: Polyclonal (identified in the batch field of the product label)

Ig isotype: Rabbit IgG

Species reactivity: In vitro diagnostics in humans. Not tested in other species

DESCRIPTION AND APPLICATION

Anti-IgM antibody reacts with surface immunoglobulin IgM mu chains. IgM is one of the predominant surface immunoglobulins on B-lymphocytes. This antibody is useful when identifying lymphomas, plasmacytomas, and B-cell lineage derived Hodgkin's lymphomas. Due

to the restricted expression of heavy and light chains in these diseases, demonstration of B-cell lymphomas is possible with clonal gene rearrangement studies.

IHC positive control: Lymph node, tonsil

Visualization: Cell cytoplasm

IHC recommended procedure:

- 4µm thick section should be taken on charged slides; dry overnight at 60°C
- Deparaffinise, rehydrate and HIER (heat induced epitope retrieval) – boil tissue in the Pt Module using Vitro S.A Tris-EDTA buffer pH9⁴ for 20 min at 95°C. Upon completion rinse with 3-5 changes of distilled or deionised water followed by cooling at RT for 20 min
- Endogenous peroxidase block - Blocking for 10 minutes at room temperature using peroxidase solution (ref. MAD-021540Q-125)
- Primary antibody: incubate for 20 minutes [The antibody dilution (when concentrated) and protocol may vary depending on the specimen preparation and specific application. Optimal conditions should be determined by the individual laboratory]
- For detection use Master Polymer Plus Detection System (HRP) (DAB included; ref. MAD-000237QK)
- Counterstaining with haematoxylin and final mounting of the slide

STORAGE AND STABILITY

Stored at 2-8°C. Do not freeze. Once the packaging has been opened it can be stored until the expiration date of the reagent indicated on the label. If the reagent has been stored under other conditions to those indicated in this document, the user must first check its correct performance taking into account the product warranty is no longer valid.

WARNINGS AND PRECAUTIONS

1. Avoid contact of reagents with eyes and mucous membranes. If reagents come into contact with sensitive areas, wash with copious amounts of water.
2. This product is harmful if swallowed.
3. Consult local or state authorities with regard to recommended method of disposal.
4. Avoid microbial contamination of reagents.

⁴ Ref: MAD-004070R/D

¹ These references are for presentation in vials of Low Density Polyethylene (LDPE) dropper. In case the products are used in automated stainers, a special reference is assigned as follows:
- / L: Cylindrical screw-cap vials (QD-3 / L, QD-7 / L, QD-12 / L).
- / N: Polygonal screw-cap vials (QD-3 / N, QD-7 / N, QD-12 / N).
For different presentations (references / volumes) please contact the supplier.

² For Technical specifications for VitroStainer 42, please contact your distributor.

³ For Technical specifications for MD-Stainer, please contact your distributor.



SAFETY RECOMMENDATIONS

This product is intended for laboratory professional use only. The product is NOT intended to be used as a drug or for domestic purposes. The current version of the Safety Data Sheet for this product can be downloaded by searching the reference number at www.vitro.bio or can be requested at regulatory@vitro.bio.

BIBLIOGRAPHY

1. Klein J, Hořejši V. Immunology. 2nd ed. Abingdon (UK): Blackwell Science Ltd; 1999. pp. 226-7, 238-46.
2. Leong ASY, Cooper K, Leong FJWA. Manual of diagnostic antibodies for immunohistology. London: Oxford University Press; pp. 217-219. 1999.
3. Swerdlow, S.H., Campo, E., Harris, N.L., Jaffe, E.S., Pileri, S.A., Stein, H., Thiele, J., Vardiman, J.W. WHO Classification of Tumours of Haematopoietic and Lymphoid Tissues, 4^a Edition, Ed. IARC, Lyon, 2008.
4. Moreira RK, Revetta F, Koehler E, Washington MK. Diagnostic utility of IgG and IgM immunohistochemistry in autoimmune liver disease. World J Gastroenterol; **16**(4): 453-457. 2010.
5. Koens L, Vermeer MH, Willemze R, Jansen PM. IgM expression on paraffin sections distinguishes primary cutaneous large B-cell lymphoma, leg type from primary cutaneous follicle center lymphoma. Am J Surg Pathol; **34**(7): 1043-1048. 2010.

LABEL AND BOX SYMBOLS

Explanation of the symbols of the product label and box:

	Expiration date
	Temperature limit
	Manufacturer
	Sufficient content for <n> assays
	Catalog number
	Lot code
	Refer to the instructions of use
	Medical product for <i>in vitro</i> diagnosis.
	Material safety data sheet

